


SEA-BIRD ELECTRONICS, INC. 1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

1808 - 136th Place Northeast, Bellevue, Washington 98005 USAPhone: (425) 643-9866Fax: (425) 643-9954Www.seabird.com

Pressure Test Certificate

<u>Customer</u>	Oregon State University
<u>Job Number</u>	47955
<u>Date</u>	9/5/2007
<u>Technician</u>	JMc
Serial Number	051980
Low Pressure (PSI)	50 PSI
<u>Time (Minutes)</u>	15 Minutes
High Pressure (PSI	10000 PSI
<u>Time (Minutes)</u>	30 Minutes
Pass 🗸	
Fail	
Comments	
Replaced pump thrust washers and "O"-rings.	
	High pressure is generally equal to the maximum depth rating of the instrument.

Pressure



Typical Test Profile