



Certificate of Calibration

Customer: Ocean Technology Group
Asset Serial Number: 020020
Asset Product Type: Smart SV&T Instrument, 500m Housing
Calibration Type: Sound Velocity
Calibration Range: 1400 to 1600 m/s
Calibration RMS Error: .0164
Calibration ID: 020020 999999 999999 080914 213225
Installed On:

Coefficient A: 7.217876E-4	Coefficient H: 0.000000E+0
Coefficient B: -7.467470E-5	Coefficient I: 0.000000E+0
Coefficient C: 9.234094E-7	Coefficient J: 0.000000E+0
Coefficient D: -2.841291E-7	Coefficient K: 0.000000E+0
Coefficient E: 0.000000E+0	Coefficient L: 0.000000E+0
Coefficient F: 0.000000E+0	Coefficient M: 0.000000E+0
Coefficient G: 0.000000E+0	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 8/9/2014

Certified By:

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at www.AMLoceanographic.com/support

AML Oceanographic
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA
T: +1-250-656-0771 F: +1-250-655-3655 Email: service@AMLoceanographic.com





Certificate of Calibration

Customer: Ocean Technology Group
Asset Serial Number: 020020
Asset Product Type: Smart SV&T Instrument, 500m Housing
Calibration Type: Temperature
Calibration Range: -2 to +32 Dec C
Calibration RMS Error: .0016
Calibration ID: 020020 888888 888888 090914 203840
Installed On:

Coefficient A: -8.301748E+0	Coefficient H: 0.000000E+0
Coefficient B: 8.901850E-4	Coefficient I: 0.000000E+0
Coefficient C: -9.846709E-9	Coefficient J: 0.000000E+0
Coefficient D: 2.169456E-13	Coefficient K: 0.000000E+0
Coefficient E: -3.037939E-18	Coefficient L: 0.000000E+0
Coefficient F: 2.783181E-23	Coefficient M: 0.000000E+0
Coefficient G: -9.545798E-29	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 9/9/2014

Certified By:

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at www.AMLoceanographic.com/support

AML Oceanographic
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA
T: +1-250-656-0771 F: +1-250-655-3655 Email: service@AMLoceanographic.com