



## Certificate of Calibration

Customer: Ocean Technology Group  
Asset Serial Number: 020020  
Asset Product Type: Smart SV&T Instrument, 500m Housing  
Calibration Type: Sound Velocity  
Calibration Range: 1400 to 1600 m/s  
Calibration RMS Error: .011  
Calibration ID: 020020 003843 204181 270216 001546  
Installed On: 020020

---

Coefficient A: 7.216886E-4	Coefficient H: 0.000000E+0
Coefficient B: -7.436780E-5	Coefficient I: 0.000000E+0
Coefficient C: 1.139299E-6	Coefficient J: 0.000000E+0
Coefficient D: -7.125645E-7	Coefficient K: 0.000000E+0
Coefficient E: 0.000000E+0	Coefficient L: 0.000000E+0
Coefficient F: 0.000000E+0	Coefficient M: 0.000000E+0
Coefficient G: 0.000000E+0	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 27/2/2016

Certified By:

Robert Haydock  
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at [www.AMLoceanographic.com/support](http://www.AMLoceanographic.com/support)

AML Oceanographic  
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA  
T: +1-250-656-0771 F: +1-250-655-3655 Email: [service@AMLoceanographic.com](mailto:service@AMLoceanographic.com)



## Certificate of Calibration

Customer: Ocean Technology Group  
Asset Serial Number: 020020  
Asset Product Type: Smart SV&T Instrument, 500m Housing  
Calibration Type: Temperature  
Calibration Range: -2 to +32 Dec C  
Calibration RMS Error: .0012  
Calibration ID: 020020 999999 020020 250216 193333  
Installed On: 020020

---

Coefficient A: -8.271178E+0	Coefficient H: 0.000000E+0
Coefficient B: 8.840404E-4	Coefficient I: 0.000000E+0
Coefficient C: -9.215226E-9	Coefficient J: 0.000000E+0
Coefficient D: 1.849194E-13	Coefficient K: 0.000000E+0
Coefficient E: -2.212972E-18	Coefficient L: 0.000000E+0
Coefficient F: 1.743399E-23	Coefficient M: 0.000000E+0
Coefficient G: -4.450266E-29	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 25/2/2016

Certified By:

Robert Haydock  
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at [www.AMLoceanographic.com/support](http://www.AMLoceanographic.com/support)

AML Oceanographic  
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA  
T: +1-250-656-0771 F: +1-250-655-3655 Email: [service@AMLoceanographic.com](mailto:service@AMLoceanographic.com)