



Certificate of Calibration

Customer: Ocean Technology Group
Asset Serial Number: 004599
Asset Product Type: Smart SV&T Instrument, 500m Housing
Calibration Type: Sound Velocity
Calibration Range: 1400 to 1550 m/s
Calibration RMS Error: .0163
Calibration ID: 004599 999999 202395 180615 222956
Installed On:

Coefficient A: 1.539509E+3	Coefficient H: 0.000000E+0
Coefficient B: -1.131910E+2	Coefficient I: 0.000000E+0
Coefficient C: 8.540824E+0	Coefficient J: 0.000000E+0
Coefficient D: -4.633874E-1	Coefficient K: 0.000000E+0
Coefficient E: 0.000000E+0	Coefficient L: 0.000000E+0
Coefficient F: 0.000000E+0	Coefficient M: 0.000000E+0
Coefficient G: 0.000000E+0	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 18/6/2015

Certified By:

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at www.AMLoceanographic.com/support

AML Oceanographic
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA
T: +1-250-656-0771 F: +1-250-655-3655 Email: service@AMLoceanographic.com





Certificate of Calibration

Customer: Ocean Technology Group
Asset Serial Number: 004599
Asset Product Type: Smart SV&T Instrument, 500m Housing
Calibration Type: Temperature
Calibration Range: -2 to +45 Deg C
Calibration RMS Error: .0275
Calibration ID: 004599 999999 T21006 170615 214741
Installed On:

Coefficient A: -4.905647E+1	Coefficient H: 0.000000E+0
Coefficient B: 3.170410E-3	Coefficient I: 0.000000E+0
Coefficient C: -5.532417E-8	Coefficient J: 0.000000E+0
Coefficient D: 5.630989E-13	Coefficient K: 0.000000E+0
Coefficient E: 0.000000E+0	Coefficient L: 0.000000E+0
Coefficient F: 0.000000E+0	Coefficient M: 0.000000E+0
Coefficient G: 0.000000E+0	Coefficient N: 0.000000E+0

Calibration Date (dd/mm/yyyy): 17/6/2015

Certified By:

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. Please note that Xchange™ sensor-heads may be installed on assets other than the one listed above; this calibration certificate will still be valid when used on other such assets. If this instrument or sensor has been recalibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary. Older generation instruments may require configuration files, which are available for download at our Customer Centre at www.AMLoceanographic.com/support

AML Oceanographic
2071 Malaview Avenue, Sidney B.C. V8L 5X6 CANADA
T: +1-250-656-0771 F: +1-250-655-3655 Email: service@AMLoceanographic.com

