

OPTICAL SCIENTIFIC, INC.

Optical Precipitation Sensor (ORG-815-Dx)

CERTIFICATE OF CALIBRATION

Date: 8 January 2009 OSi Ref No. S480166
Customer: Ocean Tech Cust. PO No. Z882904
P/N: ORG 815 DR S/N: 09010291

Optical Scientific hereby certifies compliance of the above referenced item(s) as follows:

Quality Standard: ISO-9001

OSi Test Standard: ORG9903, ORG9905, ORG0007

CALIBRATION DATA

Free Run Output (<70 mV)	<u>39</u> mV
ORG Calibration with Factory TST	<u>3372</u> mm/hr
Power Consumption (*)	<u>409</u> ma

(*) < 65 ma for ORG-815-DA with lens heaters off
< 800 ma for ORG-815-DR/DRM at 20 deg C with lens heaters on


Authorized Q.A. Representative

01/22/02

OSi is an ISO-9001 Registered Company