

OPTICAL SCIENTIFIC, INC.

Optical Precipitation Sensor (ORG-815-Dx)

CERTIFICATE OF CALIBRATION

Date: 3/23/10 OSi Ref No. S410031
Customer: Kuhio Vellales Cust. PO No. 2943436
P/N: ORG 815-DR S/N: 09100291

Optical Scientific hereby certifies compliance of the above referenced item(s) as follows:


Quality Standard: ISO-9001

OSi Test Standard: ORG9903, ORG9905, ORG0007

CALIBRATION DATA

Free Run Output (<70 mV) 40 mV
ORG Calibration with Factory TST 3.69 mm/hr
Power Consumption (*) 386 ma

(*) < 40 ma for ORG-815-DA with lens heaters off
< 800 ma for ORG-815-DR/DRM at 20 deg C with lens heaters on


Authorized Q.A. Representative

11/17/00

OSi is an ISO-9001 Registered Company