



Optical Scientific Inc.

**OPTICAL SCIENTIFIC, INC.**

Optical Precipitation Sensor (ORG-815-DXXX)

**CERTIFICATE OF CALIBRATION**

Date: 26 December 2013 OSi Ref No. 5414058  
Customer: Ocean Technology Cust. PO No. 210065428  
P/N: ORG 815 DR S/N: 09100291

Optical Scientific hereby certifies compliance of the above referenced item(s) as follows:

Quality Standard: ISO-9001

OSi Test Standard: ORG9903, ORG9905, ORG0007

**CALIBRATION DATA**

Free Run Output (40 +/- 5mV)	<u>40</u> mV
ORG Calibration (Voltage)	<u>3.359</u> Vdc
Power Consumption (<800)	<u>457</u> ma

  
Authorized Q.A. Representative

05/11/10

OSi is an ISO-9001 Registered Company