



Optical Scientific Inc.

OPTICAL SCIENTIFIC, INC.

Optical Precipitation Sensor (ORG-815-DXXX)

CERTIFICATE OF CALIBRATION

Date: 26 March 2012 OSi Ref No. 5412028
Customer: Ocean Technology Cust. PO No. Z A15124
P/N: ORG 815 DR S/N: 09050297


Optical Scientific hereby certifies compliance of the above referenced item(s) as follows:

Quality Standard: ISO-9001

OSi Test Standard: ORG9903, ORG9905, ORG0007

CALIBRATION DATA

Free Run Output (40 +/- 5mV)	<u>40</u> mV
ORG Calibration (Voltage)	<u>4.26</u> Vdc
Power Consumption (<800)	<u>435</u> ma


Authorized Q.A. Representative

05/11/10

OSi is an ISO-9001 Registered Company