

OPTICAL SCIENTIFIC, INC.

Optical Precipitation Sensor (ORG-815-Dx)

CERTIFICATE OF CALIBRATION

Date: 5/29/09 OSi Ref No. 5490066
Customer: Victor Polidoro Cust. PO No. 2911518
P/N: ORG 815DR S/N: 09050297

Optical Scientific hereby certifies compliance of the above referenced item(s) as follows:


Quality Standard: ISO-9001

OSi Test Standard: ORG9903, ORG9905, ORG0007

CALIBRATION DATA

Free Run Output (<70 mV) 40 mV
ORG Calibration with Factory TST 3.652 mm/hr
Power Consumption (*) 478 ma

(*) < 65 ma for ORG-815-DA with lens heaters off
< 800 ma for ORG-815-DR/DRM at 20 deg C with lens heaters on


Authorized Q.A. Representative

01/22/02

OSi is an ISO-9001 Registered Company