



Optical Scientific Inc.

TST-800 ORG Field Test Kit
CERTIFICATE OF CALIBRATION

DATE : <u>14 January 2008</u>	OSi SO # :	<u>5480166</u>
CUST : <u>Ocean</u>	CUST REF # :	<u>Z882904</u>
P/N: <u>TST-800</u>	S/N:	<u>09010054</u>

Optical Scientific Inc (OSi) hereby certifies compliance on the above referenced item as follows; Quantities are those specified by our packing list. All materials used are new. All technical requirements of the customer order are satisfied including these pertaining to material and processes. Quality and performance are in accordance with customer order terms, conditions and instructions, specifications, and any references and associated contractual drawings and documents.

CALIBRATION DATA

FREQUENCY OUTPUT: 50.0 Hz (50 +/- 1Hz)

CARRIER FREQUENCY: 46.0 KHz (46 +/- 1KHz)

CALIBRATION VOLTAGE : 3.784 Vdc (3.784 +/- .001Vdc)

- AT STANDARD LAB CONDITIONS, 20 deg C and 50% RH USING OSi LAB STANDARDS
- LAB STANDARDS; ORG/TST 800 SERIAL #001, ORG 815 SERIAL # 001
- QUALITY STANDARD: ISO 9001

AUTHORIZED QA/QC REPRESENTATIVE: [Signature]