

Calibration Record

Date: Oct. 21, 2010
Sensor Type: Seapoint Chlorophyll Fluorometer
Serial Number(s): 2440

A comparative calibration was performed using a calibrated reference fluorometer. The reference fluorometer was calibrated with the cultured algae *Isochrysis galbana*.

The Gain B switch will not set low, so only 3X or 30X gains function.

Aside from the gain switch, this Seapoint Chlorophyll Fluorometer meets or exceeds the specifications stated in the supplied User Manual.

If you have any questions please contact me at 603/642-4921 or seapoint@seapoint.com

T. G. Burke

11/22/10

Signature

Date



PO Box 368 • Exeter, NH 03833 • USA
Tel: (603) 642-4921 • Fax: (603) 642-4922
seapoint@seapoint.com • www.seapoint.com