

## Calibration Record

Date: January 21, 2009  
Sensor Type: Seapoint Chlorophyll Fluorometer  
Serial Number(s): 2487

A comparative calibration was performed using a calibrated reference fluorometer. The reference fluorometer was calibrated with the cultured algae *Isochrysis galbana*.

This Seapoint Chlorophyll Fluorometer meets or exceeds the specifications stated in the supplied User Manual.

If you have any questions please contact me at 603/642-4921 or [seapoint@seapoint.com](mailto:seapoint@seapoint.com)

*T. G. Blake*

1/21/09

---

Signature

Date

*Received  
1/29/09  
JM*



PO Box 368 • Exeter, NH 03833 • USA  
Tel: (603) 642-4921 • Fax: (603) 642-4922  
[seapoint@seapoint.com](mailto:seapoint@seapoint.com) • [www.seapoint.com](http://www.seapoint.com)