


## Calibration Record

Date: April 28, 2011  
Sensor Type: Seapoint Chlorophyll Fluorometer  
Serial Number(s): 2487

A comparative calibration was performed using a calibrated reference fluorometer. The reference fluorometer was calibrated with the cultured algae *Isochrysis galbana*.

This Seapoint Chlorophyll Fluorometer meets or exceeds the specifications stated in the supplied User Manual.

If you have any questions please contact me at 603/642-4921 or [seapoint@seapoint.com](mailto:seapoint@seapoint.com)

  
\_\_\_\_\_  
Signature

4/28/11  
\_\_\_\_\_  
Date



PO Box 368 • Exeter, NH 03833 • USA  
Tel: (603) 642-4921 • Fax: (603) 642-4922  
[seapoint@seapoint.com](mailto:seapoint@seapoint.com) • [www.seapoint.com](http://www.seapoint.com)