



CALIBRATION CERTIFICATE

Before adjustment

Instrument PTB220 Class A Digital Barometer
Serial number C2750001
Manufacturer Vaisala Oyj, Finland
Calibration date 09th March 2012

The above instrument was calibrated by comparing the readings of the instrument to the factory working standard of Vaisala.

The pressure readings of the factory working standard have been calibrated at an ISO/IEC 17025 accredited calibration laboratory (FINAS), Vaisala Measurement Standards Laboratory (MSL), by using MSL working standards traceable to NIST.

Calibration results

Reference hPa	Observed hPa	Correction* hPa
500.00	500.02	-0.02
550.02	550.03	-0.01
650.02	650.02	0.00
750.01	750.01	0.00
850.01	850.02	-0.01
950.02	950.03	-0.01
1000.01	1000.03	-0.02
1050.01	1050.03	-0.02
1100.01	1100.03	-0.02

*To obtain the true pressure, add the correction to the barometer reading.
Interpolated corrections may be used at intermediate readings of the scale of the barometer.

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
PPC4	439	2011-10-21	3336095-439:1319195158

Uncertainty (95 % confidence level, k=2)

Pressure ± 0.07 hPa

Ambient Conditions

Humidity 34 %RH \pm 5 %RH
Temperature 22 °C \pm 1 °C
Pressure 1002-hPa \pm 1 hPa

Approved By

Technical Operator



CALIBRATION CERTIFICATE

After adjustment

Instrument PTB220 Class A Digital Barometer
Serial number C2750001
Manufacturer Vaisala Oyj, Finland
Calibration date 09th March 2012

The above instrument was calibrated by comparing the readings of the instrument to the factory working standard of Vaisala.

The pressure readings of the factory working standard have been calibrated at an ISO/IEC 17025 accredited calibration laboratory (FINAS), Vaisala Measurement Standards Laboratory (MSL), by using MSL working standards traceable to NIST.

Calibration results

Reference hPa	Observed hPa	Correction* hPa	Acceptance limit hPa
500.01	500.02	-0.01	± 0.05
550.01	550.01	0.00	± 0.05
650.00	650.00	0.00	± 0.05
750.02	750.02	0.00	± 0.05
850.00	850.00	0.00	± 0.05
950.01	950.01	0.00	± 0.05
1000.01	1000.00	0.01	± 0.05
1050.01	1050.01	0.00	± 0.05
1100.01	1100.01	0.00	± 0.05

*To obtain the true pressure, add the correction to the barometer reading.
 Interpolated corrections may be used at intermediate readings of the scale of the barometer.

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
PPC4	439	2011-10-21	3336095-439:1319195158

Uncertainty (95 % confidence level, k=2)


Pressure ± 0.07 hPa

Ambient Conditions

Humidity 33 %RH ± 5 %RH
 Temperature 23 °C ± 1 °C
 Pressure 1007 hPa ± 1 hPa



 Approved By



 Technical Operator