PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

20.3 °C

C-Star Calibration

| Date | April 7, 2011 | S/N# | CST-1192DR | Pathlength 25 cm |
|----------------------------------|---------------|------|--------------|------------------|
| | | | Analog meter | |
| V_d | | | 0.060 V | |
| V_{air}^{u} | | | 4.821 V | |
| V_{ref} | | | 4.723 V | |
| Temperature of calibration water | | | | 21.6 °C |

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: c = -1/x * In (Tr)

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref}.

Ambient temperature: meter temperature in air during the calibration.

 V_{sig} Measured signal output of meter.

Ambient temperature during calibration

Revision L 6/9/09