



Chlorophyll WETStar Characterization

Date: March 24, 2011

S/N: WS3S-1352

Chlorophyll concentration expressed in $\mu\text{g/l}$ can be derived using the equation:

$$\text{CHL}(\mu\text{g/l}) = \text{Scale Factor} \times (\text{Output} - \text{Clean Water Offset})$$

	Analog output	Digital output
Clean Water Offset (CWO)	0.084 V @	86 counts
Scale Factor (SF)	15.3 $\mu\text{g/l/V}$ @	0.0184 $\mu\text{g/l/count}$
Maximum Output	5.48 V @	4095 counts
Resolution	0.53 mV	1 counts
Ambient Characterization Temperature	22 \pm 1 $^{\circ}\text{C}$	
Current Draw	0 mA @ 12V (typical)	
12-hour Stability	0.47 mV/hr	1 counts/hr
Temperature Stability, 25–2 $^{\circ}\text{C}$	0.12 mV/ $^{\circ}\text{C}$	1 counts/ $^{\circ}\text{C}$

Definitions:

CWO: Clean Water Offset value obtained using pure filtered de-ionized water.

SF: Scale Factor is used to convert the fluorescence response of the instrument into chlorophyll-a concentration. Scale Factor is determined at WET Labs during a cross calibration using a solid fluorescent standard and a reference fluorometer whose chlorophyll fluorescence response has been characterized in a laboratory using a mono-species lab culture of *Thalassiosira weissflogii* phytoplankton.

Maximum Output: Maximum signal output of the fluorometer.

Resolution: Standard deviation of 1 minute of clean water data, sampled once per second.

Ambient Characterization Temperature: Room temperature at time of characterization.

Current Draw: The amount of current the instrument uses for operation.

12-hour Stability: Deviation of output averaged over 12 hours.

Temperature Stability: Measured output variation per degree.