



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0273  
 CALIBRATION DATE: 09-Mar-19

Glider APL CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00614071e+001      CPcor = -9.5700e-008 (nominal)  
 h = 1.13344446e+000      CTcor = 3.2500e-006 (nominal)  
 i = -3.01291631e-003  
 j = 2.96551151e-004

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.98779	0.00000	0.00000
1.0000	34.6996	2.96694	5.94009	2.96695	0.00001
4.5000	34.6803	3.27315	6.16442	3.27315	-0.00001
15.0000	34.6391	4.25217	6.83167	4.25217	-0.00000
18.4999	34.6300	4.59630	7.05096	4.59630	-0.00000
23.9999	34.6184	5.15242	7.39128	5.15243	0.00001
28.9998	34.6080	5.67201	7.69530	5.67201	-0.00000
32.4999	34.5955	6.04182	7.90435	6.04181	-0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

$$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$$

