

Sea-Bird Electronics, Inc.

13431 NE 20th Street Bellevue, WA 98005 United States

Serv

Phone +1-425-643-9866 Fax +1-425-643-9954 www.seabird.com

 Service Request
 1005506442

 Date
 14-MAR-2019

 Sales Order
 315434754

PO Number:

CUSTOMER INFORMATION

SERVICE REPORT

Name: KONGSBERG UNDERWATER TECHNOLOGY INC

Account: 40280537

ROSS HEALY

ross.healy@km.kongsberg-us.com

2069538309

Bill To Address

19210 33RD AVE W;SUITE A; LYNNWOOD,WA,98036,US **Ship To Address**

19210 33RD AVE W;SUITE A; LYNNWOOD,WA,98036,US

PRODUCT INFORMATION

Item: APL-GLIDER.LEGACY

Item Description: (LEGACY) APL Glider

Serial: 0273

Special Notes

Services Requested:

Evaluate/Repair Instrumentation.
Perform Routine Calibration Service.
Replaced the conductivity cell.
Replaced the electronic board sets.

Services Performed:

Perform initial diagnostic evaluation. *
Replaced the electronic board set.
Replaced the temperature and conductivty boards.
Performed pressure calibration.
Performed a "Final" Calibration.

^{*} Customer returned the cell guard but not the stanchion with the order.

Item	Item Description	Qty
41119T	UW/APL GLIDER ISOLATED WBO COND ELECT, TESTED	1
41028T	UW/APL GLIDER WBO TEMP ELEC'S ASSY, TESTED	1
REPGLIDERCELL	Extra charge to replace broken SBE Glider conductivity cell	1
CAL_SEAGLIDER	Calibrate Seaglider conductivity and temperature sensors	1
CNCRTSEAGLDR	CONFIRM & RE-CERTIFY SEAGLIDER CTD	1

Unbilled Items

Item	Item Description	Qty
802761	ALACE COND CELL ASSEMBLY, NO WRAP, SHIELDED WIRES	1



Sea-Bird Electronics, Inc.

13431 NE 20th Street Bellevue, WA 98005 United States

SERVICE REPORT

Phone +1-425-643-9866 Fax +1-425-643-9954 www.seabird.com

 Service Request
 1005506442

 Date
 14-MAR-2019

 Sales Order
 315434754