



## Sea-Bird Electronics, Inc.

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: 01/19/16

Description: SBE-43F DO Sensor

### Sensor Information:

Model Number: 43F

Serial Number: 3327

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 870 PSI Held For: 15 Minutes

Passed Test: Yes

Tested By: RH

High pressure is generally equal to the maximum depth rating of the instrument

